



85 105

	Jet 1	Jet 2	
Speech Understanding	Multiband Adaptive Directionality	•	•
	Noise Reduction	•	•
	Single Compression	•	•
	Frequency lowering	Speech Rescue™	-
Sound Quality	Fitting Bandwidth*	8 kHz	8 kHz
	Processing Channels	48	48
Listening Comfort	Feedback Management	Feedback shield LX	Feedback shield LX
	Transient Noise Management	On/Off	-
	Wind Noise Management	•	•
Personalisation & Optimising Fitting	Fitting Bands	10	8
	Fitting Formulas	NAL-NL1/NAL-NL2, DSL v5.0	NAL-NL1/NAL-NL2, DSL v5.0
Connecting to the world	Direct streaming**	•	•
	Oticon ON app & Oticon RemoteCare app	•	•
	ConnectClip	•	•
	EduMic	•	•
	Remote Control 3.0	•	•
	TV Adapter 3.0	•	•
	Phone Adapter 2.0	•	•
Tinnitus SoundSupport™	•	•	

* Bandwidth accessible for gain adjustments during fitting

** From compatible iPhone®, iPad®, iPod touch® devices

Operating Conditions

Temperature: +1°C to +40°C (34°F to 104°F)
 Humidity: 5% to 93% relative humidity, non-condensing
 Atmospheric pressure: 700 hPa to 1060 hPa

Storage and transportation conditions

Temperature and humidity should not exceed the below limits for extended periods during transportation and storage.

Transportation

Temperature: -25°C to +60°C (-13°F to 140°F)
 Humidity: 5% to 93% relative humidity, non-condensing
 Atmospheric pressure: 700 hPa to 1060 hPa

Storage









Temperature: -25°C to +60°C (-13°F to 140°F)
 Humidity: 5% to 93% relative humidity, non-condensing
 Atmospheric pressure: 700 hPa to 1060 hPa

Apple, the Apple logo, iPhone, iPad, and iPod touch are trademarks of Apple Inc., registered in the U.S. and other countries.











IP68

For information on compatibility, please visit www.oticon.com.au/compatibility

		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+ AMD1:1998 CSV and IEC 60318-4:2010	2CC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006
 <p>85</p>  <p>Hook</p> <p>Corda minifit</p> <p>Technical information Omnidirectional mode is used unless otherwise stated.</p>		OSPL90  <p>Full-on gain</p>  <p>Frequency Response</p> 	OSPL90  <p>Full-on gain</p>  <p>Frequency Response</p> 
		Peak OSPL90 130 (125 ¹) dB SPL 1600 Hz 121 (107 ¹) dB SPL HFA-OSPL90 122 (113 ¹) dB SPL	Peak OSPL90 120 (121 ¹) dB SPL 1600 Hz 113 (98 ¹) dB SPL HFA-OSPL90 115 (105 ¹) dB SPL
	Peak Full-on gain ² 66 (59 ¹) dB 1600 Hz 55 (41 ¹) dB HFA-FOG 57 (49 ¹) dB	Peak Full-on gain ² 57 (54 ¹) dB 1600 Hz 47 (33 ¹) dB HFA-FOG 50 (41 ¹) dB	
	Reference test gain	46 dB	39 dB
	Frequency range	105-7500 Hz	100-7000 Hz
	Telecoil output (1600 Hz)	1 mA/m field 85 dB SPL 10 mA/m field 105 dB SPL SPLITS L/R -	- - 97/97 dB SPL
	Total harmonic distortion (Input 70 dB SPL)	500 Hz <2 % 800 Hz 2 % 1600 Hz <2 %	<2 % <2 % <2 %
	Equivalent input noise level	Omni 21 dB SPL Dir 31 dB SPL	18 dB SPL 28 dB SPL
	Battery consumption ³	Typical 1.4 mA Quiescent 1.3 mA	1.7 mA 1.7 mA
	Battery life, artificial measurement, hours ⁴	230	180
	Expected battery life, hours (battery size 13 - IEC PR48) ⁵	105 - 115	

1) For instruments fitted with Corda miniFit
 2) Measured with the gain control of the hearing aids set to their full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0:1983+A1:1994 but without influence of feedback.
 3) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.
 4) Based on the standardised battery consumption measurement (e.g. IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.
 5) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

		Ear Simulator Measured according to IEC 60118-0:1983/AMD1:1994, IEC 60118-0:2015, IEC 60118-1:1995+ AMD1:1998 CSV and IEC 60318-4:2010	2CC Coupler Measured according to ANSI S3.22-2014, IEC 60118-0:2015 and IEC 60318-5:2006
 <p>105</p> <p>Hook</p> <p>Corda minifit</p>		OSPL90 	OSPL90 
		Full-on gain 	Full-on gain 
		Frequency Response 	Frequency Response 
OSPL90	Peak	138 (132 ¹) dB SPL	131 (129 ¹) dB SPL
	1600 Hz	130 (121 ¹) dB SPL	123 (113 ¹) dB SPL
	HFA-OSPL90	133 (126 ¹) dB SPL	126 (118 ¹) dB SPL
Full-on gain ²	Peak	73 (69 ¹) dB	66 (66 ¹) dB
	1600 Hz	65 (56 ¹) dB	57 (47 ¹) dB
	HFA-FOG	68 (62 ¹) dB	61 (54 ¹) dB
Reference test gain		57 dB	50 dB
Frequency range		150-7300 Hz	120-7000 Hz
Telecoil output (1600 Hz)	1 mA/m field	97 dB SPL	-
	10 mA/m field	117 dB SPL	-
	SPLITS L/R	-	109/109 dB SPL
Total harmonic distortion (Input 70 dB SPL)	500 Hz	7 %	3 %
	800 Hz	5 %	<2 %
	1600 Hz	<2 %	<2 %
Equivalent input noise level	Omni	17 dB SPL	14 dB SPL
	Dir	29 dB SPL	27 dB SPL
Battery consumption ³	Typical	1.8 mA	1.9 mA
	Quiescent	1.6 mA	1.6 mA
Battery life, artificial measurement, hours ⁴		175	160
Expected battery life, hours (battery size 13 - IEC PR48) ⁵		80-105	

1) For instruments fitted with Corda miniFit Power.
 2) Measured with the gain control of the hearing aids set to their full-on position minus 20 dB and with an input SPL of 70 dB. This is to obtain a gain response equal to the full-on gain response from e.g. IEC 60118-0:1983+A1:1994 but without influence of feedback.
 3) Battery current is measured according to IEC 60118-0:1983/AMD1:1994 §7.11, IEC 60118-0:2015 §7.7 and ANSI S3.22:2014 §6.13 after a settling time of minimum 3 minutes.
 4) Based on the standardised battery consumption measurement (e.g. IEC 60118-0:1983/AMD1:1994). The actual battery life depends on battery quality, use pattern, active feature set, hearing loss and sound environment.
 5) Real usage battery life is shown as an estimated interval based on mixed use cases with variable amplification settings and variable input levels, incl. direct stereo streaming from a TV (25% of the time) and streaming from a mobile phone (6% of the time).

Headquarters
Oticon A/S
Kongebakken 9
DK-2765 Smørum
Denmark

SBO Hearing A/S
Kongebakken 9
DK-2765 Smørum
Denmark