Instructions for use



Streamer Pro 1.3A

ConnectLine







Introduction to this booklet

This booklet guides you on how to use and maintain your new Streamer. Please read the booklet carefully, including the Warning section. This will help you achieve the full benefit of your new Streamer.

If you have additional questions about the use of your Streamer, please contact your hearing care professional.

The Streamer connects your hearing aids wirelessly to different audio sources and makes your hearing aids work like wireless headphones. The Streamer transmits the sound directly into both hearing aids and thereby improves the audio experience.

The complete system description is available online at www.oticon.global/connectline. Here you will find interactive presentations, "how-to videos" and a support section with frequently asked questions.

Intended use

Streamer Pro 1.3A is intended to be used by wireless hearing solutions users. The product is intended to be used together with the given hearing solution. The streamer is a communication interface between the hearing solutions and other devices such as cell phones. The Streamer can be connected to a range of external (additional) devices to enable the hearing instrument user to better hear a TV or to communicate through Bluetooth-enabled cell phones, etc.

The Streamer can also be used as a remote control that enables you to adjust the hearing instrument volume and change programs in the hearing instruments.

The Streamer is not intended to be used by infants/children < 36 months.

"Made for iPod," "Made for iPhone," and "Made for iPad" mean that an electronic accessory has been designed to connect specifically to iPod, iPhone, or iPad, respectively, and has been certified by the developer to meet Apple performance standards. Apple is not responsible for the operation of this device or its compliance with safety and regulatory standards. Please note that the use of this accessory with iPod, iPhone, or iPad may affect wireless performance.

Apple, the Apple logo, iPhone, iPad, and iPod touch are trademarks of Apple Inc., registered in the U.S. and other countries. App Store is a service mark of Apple Inc. Android, Google Play, and the Google Play logo are trademarks of Google LLC. The Android robot is reproduced or modified from work created and shared by Google and used according to terms described in the Creative Commons 3.0 Attribution License.

Contents

Package content	
Streamer overview	1
Basic functions - first-time use	
Wearing the Streamer with the neck loop	1
Using the Streamer without the neck loop	1
Turning the Streamer on and off	1
Charging the Streamer	1
Battery low indicator	2
Remote control of the hearing aids	2
Pairing the Streamer with a mobile phone and other devices	2
Clear pairing	2
Primary functions	
Connectivity/streaming	3
Mobile phone	3

Continues on next page

ConnectLine phone solution	4
ConnectLine microphone	4
ConnectLine TV solution	4
Secondary functions	
ConnectLine App	4
Connectivity/streaming	4
Music/audio cable	4
Headset cable	ļ
Music/audio wireless	
PC headset - wireless	ļ
T-coil (built-in)	į
FM receiver (optional)	į

Additional information	
Cleaning and maintenance	
Accessories and spare parts	
Warnings	
Technical information	
International warranty	

Package content

The Streamer box consists of the following elements:





,-----


Protective skin (page 62)

8


Box 3












This guide (Instructions for use)



IMPORTANT NOTICE

Please refer to your hearing aid's Instructions for use for details on the functionality of your specific hearing aid solution.

10



Basic functions - first-time use

Wearing the Streamer with the neck loop

The neck loop is an integral part of the Streamer's electronic system. It is used as an antenna to transmit sound to your hearing aids.

For the best possible sound and longest possible streaming time, the Streamer should always be worn with one of the included neck loops. The Streamer comes with two neck loops, a long and a medium.


IMPORTANT NOTICE


Please use the neck loop carefully. To avoid strangulation the neck loop will automatically release if it gets caught in something. Do not shorten or modify the neck loop in any way, such as tie a knot to shorten the length. If a neck loop breaks, it cannot be repaired and must be replaced immediately. Consult your hearing care professional for a replacement.

Inserting neck loop

The neck loop snaps into the holes on the top of the Streamer.

Note that when talking to somebody through the Streamer, it is important that your voice can be picked up by the built-in microphone on the top.






Using the Streamer without the neck loop

When using the Streamer without the neck loop, the distance to the hearing aid should be less than one metre. You may experience drop-outs in sound, which is why we always recommend using the neck loop.

When using the Streamer without the neck loop, more power will be used - consequently maximum streaming time is reduced to 5 hours on one full charge.



Turning the Streamer on and off

Power on

Press the on/off button on the top of the Streamer for approx. 2 seconds until all lights on the front and top flash to indicate it is powered on.

Release the on/off button. The power indicator on the top of the Streamer should now fade to green.


If the Streamer cannot power on, make sure it is charged (page 18).

To see the status of the Streamer, press the on/off button briefly.

Power off

Press and hold the on/off button on the top of the Streamer for approx. 2 seconds until the green power indicator is turned off.

16



Power on/off		
Status	Status Front Description	
000		The Streamer is off
•••	/	After pressing the on/off button for approx. 2 seconds, all lights will be on for a few seconds to indicate that the Streamer is now turned on
•••	P	The Streamer is now turned on and ready for use

Charging the Streamer

The Streamer uses a rechargeable battery. We recommend that you charge the Streamer once a day for best possible performance.

Insert cable from the enclosed charger in the micro USB socket in the bottom of the Streamer (A). Remember to connect the charger to a power outlet.

Charging, when turned on, is indicated by a blinking green indicator on the top. Once fully charged, the indicator will show steady green.

Charging, when turned off, is indicated by a blinking green indicator on the top. Once fully charged, the indicator will be off.


On/off indicator

IMPORTANT NOTICE

When recharging the Streamer use the supplied original power supply 5.0V/1A, CE and safety marked. $\square \uparrow \uparrow$

Charging		
Status Front Description		
		Charging - when turned off - status indicator will blink green
		Charging - when turned on (or in use) - the status indicator will shift between strong and soft green

□ Blink □ Fade




IMPORTANT NOTICE

When recharging, the on/off indicator should be lit as described in this Instructions for use. If the on/off indicator is not lit, disconnect the Streamer from the charger and consult your hearing care professional.

Battery low indicator

The Streamer will indicate battery low (approx. 20 minutes capacity left) with a steady red light, and battery very low (approx. 5 minutes capacity left) with a blinking red light.



Battery low indicator		
Status	Front	Description
		Battery low - approx. 20 minutes capacity left when streaming to hearing aids
		Battery very low – approx. 5 minutes capacity left when streaming to hearing aids

Steady Fading

IMPORTANT NOTICE

The battery should not be drained completely. Keeping the battery fully charged will prolong the capacity of the battery. The battery cannot be replaced by the user.

Battery capacity and charge time		
Battery capacity using neck loop	Up to 10 hours typical daily use: 2 hours phone, 6 hours TV and 2 hours standby.	
Battery capacity not using neck loop	Up to 5 hours typical use: 1 hour phone, 3 hours TV and 1 hour standby.	
Battery charge time	2 hours for complete charge, using enclosed power supply.	
Remote control	If the Streamer is turned off, remote control of hearing aid will work for up to 6 months on one full charge.	
Battery life expectancy	The battery in the Streamer will last for more than 2 years if fully charged once a day. The battery can be replaced by your hearing care professional.	

Remote control of the hearing aids

Volume up/down

Briefly pressing on either volume up or down will regulate the volume in the hearing aids accordingly. You will hear small confirmation "ticks" in the hearing aids.

Changing programs in the hearing aids


You can change programs up or down by holding the volume up or down for more than one second. Changing programmes is indicated by a sound in your hearing aids.

Muting the hearing aids

You can mute the hearing aids by pressing the volume up AND down for 1 second simultaneously.

IMPORTANT NOTICE

- The mute function is only available in the most recent Oticon hearing aids. Please consult your hearing care professional for details.
- Even though the Streamer is turned off, the remote functionality of the hearing aids will still work.




Pairing the Streamer with a mobile phone and other devices

Before the Streamer can be used with other devices - such as mobile phones - the two devices have to be paired.

26


1. Put the Streamer in pairing mode

Press and hold the on/off button for approx. 5 seconds, until the indicator blinks blue.



2. Put the other device (e.g. mobile) in pairing mode.


This is typically in a set-up menu on your phone. We advise you to look in your mobile phone's "instructions for use", e.q under pairing.



The pin code for the Streamer is 0 0 0 0 (four zeros).

4. Once paired, the mobile phone will typically show a headset icon; at the same time the Streamer pairing indicator will be steady blue.








Clear pairing

If you have problems with pairings, you can clear all pairings by pressing the on/off button and the volume up button at the same time for 3 seconds.



Pairing		
Status Front Description		Description
		Streamer in pairing mode
		Pairing complete
• (P) (C)		Streamer on

Steady Blinking

Static electricity

If the streamer is exposed to electrostatic discharge, i.e. static electricity, it may stop streaming. The Streamer will by default restart and you can re-activate your streaming.

For detailed information on how to establish specific streaming connections, please refer to "Primary functions" (page 32) or "Secondary functions" (page 46) in this Instructions for use.

Primary functions

Connectivity/streaming

The ConnectLine Streamer user interface has been designed to give you the best possible daily use in the following situations:

1. Use of a mobile phone through Bluetooth* (page 33).


Use of the ConnectLine phone solution (sold separately) with an analogue landline phone* (page 40).

- 2. Use of a ConnectLine partner microphone (sold separately), to improve your chances of hearing a chosen speaker at a distance, or in very noisy situations (page 42).
- Use of ConnectLine TV solution (sold separately) to wirelessly transmit the TV sound to your hearing instruments* (page 44).

* In effect turning the Streamer and the hearing aids into a wireless headset

Mobile phone

To be able to receive or make calls while using the Streamer, make sure that the Streamer and mobile phone have been paired (pages 26-27), are turned on and are within range (10 metres). Your mobile phone will often indicate that it is connected to the Streamer with a small headset icon.



How to answer a call


Make sure that you are able to speak into the Streamer's microphone.

When your mobile phone rings, you will hear an incoming call alert in your hearing aids. At the same time the phone button on the Streamer and the status indicator will blink green.

Press the phone button on the Streamer briefly to answer the call. You will hear a small confirmation beep in your hearing aids.

The phone button and status indicator will now be steady green.

To end the call, simply press the phone button on the Streamer briefly.



Receiving a call		
Status	Front	Description
• (p) ()		Streamer on and ready
		Phone indicator and top status blinks green. You will hear an incoming call indicator in your instruments
		When the call is connected, you will hear a beep acknowledging that the call is answered - you are now able to listen and talk to the other party

Steady Blinking


How to make a call

Make sure that you are able to speak into the Streamer's microphone.

To make a call, you only need to operate the mobile phone. You do not need to operate the Streamer. Simply key in the number as you normally do to make a call - the call is automatically connected to your Streamer.

The phone button on the Streamer and the status indicator on the top will turn green, and you will hear a confirmation beep in your hearing aids indicating that the call is now connected through your Streamer and hearing aids.

To end the call, simply press the "call end" button on the phone or the phone button on the Streamer briefly.





Making a call			
Status Front Description			
• (P) (C)		Streamer on and ready	
••••		When the call is connected, you will hear a beep acknowledging that the call is connected - you are able to listen and talk to the other party	

Steady Blinking

Personalisation of mobile phone features

The following mobile phone features can be enabled, or disabled, by your hearing care professional.

Reject call - default on

An incoming call can be rejected by briefly pressing the volume down button on the Streamer.

Voice dial - default on

Briefly pressing the phone button will enable voice dial (if supported by your mobile phone and you are not in range of a ConnectLine phone adapter).

Last number redial - default on

Pressing and holding the phone button will activate the last number redial (if supported by mobile phone).

Transfer call - default on

Pressing and holding the phone button will transfer a call back to the mobile phone.

All front buttons answer call - default on

If off, only the phone button on the Streamer will answer calls.

Enable incoming (A2DP) connections - default off

If enabled, other devices, e.g. iPods, computers and mobile phones can activate/start audio streaming, for example, incoming text alerts or music started from a paired and connected mobile phone.

ConnectLine phone solution


The Streamer can also be used with a traditional analogue landline phone by using a ConnectLine phone solution, which is available as a separate accessory.

For details on how to use the ConnectLine phone solution, please refer to the Instructions for use included with the ConnectLine phone solution.



Talking on the phone		
Status	Front	Description
		Streamer on and ready
••••		Streamer connected – you are able to listen and talk to the other party

Steady




ConnectLine microphone


As a separate accessory, a ConnectLine microphone is available. Speech from the person wearing the microphone is transferred through the streamer to the person wearing Oticon hearing aids.

For details on how to use the ConnectLine microphone solution, please refer to the Instructions for use included with the ConnectLine microphone.

42




Sound from ConnectLine microphone		
Status Front Description		Description
• (P) (C)		Streamer on and ready
••••		When connected to the microphone – you will hear sound from the microphone in hearing aids



ConnectLine TV solution

As a separate accessory, a ConnectLine TV solution is available, which transmits the sound from a TV to your Oticon hearing aids through the Streamer.

For details on how to use the ConnectLine TV solution, please refer to the Instructions for use included with the ConnectLine TV solution.



Watching TV			
Status	Front	Description	
• ()		Streamer on and ready	
••••		Connecting to TV	
•••		Streamer connected - sound from TV in hearing instruments	
Steady •	Steady O O Blinking		
45			

Secondary functions

ConnectLine App

The app for iPhone®, iPad®, iPod touch® and Android™ devices offers an intuitive and discreet way to control the hearing aids and the ConnectLine system. Also, it offers possibility to customise the features of the ConnectLine system.

Available on App Store® and on Google Play™.





46


Please visit www.oticon.global/connectline for more information and for details on compatibility.



Connectivity/streaming

The Streamer can be used with a wide range of standard peripherals, such as MP3 players, computers, etc.

The following priorities apply.




Α	Priority 1	Mini jack in/out (if connected)	Briefly press AUX selector	Page 48
В	Priority 2	FM (if connected)	Briefly press AUX selector	Page 58
С	Priority 3	Wireless Bluetooth (built-in)	Briefly press AUX selector	Page 52
D	Priority 4	Tele loop receiver (built-in)	Press and hold AUX selector for approx. 2 seconds	Page 56


Music/audio cable

Inserting and connecting the included mini jack cable enables you to listen to music/audio from a wide range of devices.

To listen to audio from your device, simply press the AUX selector briefly. When the cable is inserted, streaming will start automatically.



Music/audio - cable		
Status Front Description		
• (P) (C)		Streamer on and ready
• •		You will hear a confirmation beep followed by the sound from the device



Headset cable


Make sure that you are able to speak into the Streamer's microphone.

Inserting and connecting the included mini jack cable and the included headset splitter cable enables you to use the Streamer as a headset with your computer.

To use the headset function, simply press the AUX selector briefly.



Headset cable		
Status Front Description		
• (P) (C)		Streamer on and ready
••••		You will hear a confirmation beep - you are now able to listen and talk to the other party



50 51

Steady


Music/audio wireless

To hear music/sound wirelessly, e.g. from an MP3 player or a computer with built-in Bluetooth, make sure that the device has been paired to your Streamer (page 26), that both are turned on, and within range (<10 metres).

To listen to music e.g. from a mobile phone, press the AUX selector before starting the music player.

Enable automatic connections – default off
If enabled by your hearing care professional, iPods, computers and mobile phones can automatically activate/
start audio streaming, for example, incoming text alerts or music started from a paired and connected mobile phone.

52



Music/audio wireless		
Status Front Description		
• (p) ()		Streamer on and ready
• (p) •		Streamer connecting to music/audio device
• (i) •		When connected, you will hear a confirmation beep followed by the sound from the device

Steady Blink



PC headset - wireless


Make sure that you are able to speak into the Streamer's microphone inlet.

When pairing to a computer, make sure that the Streamer is paired as a headset device.

To use the headset function, call control is done from the computer.



Headset - wireless		
Status Front Description		
• (p) ()		Streamer on and ready
• (p) •		Streamer connecting to music/audio device
• (i) •		When connected, you will hear a confirmation beep followed by the sound from the device




T-coil (built-in)

The Streamer has a built-in tele loop receiver (T-coil).

To receive sound from the tele loop receiver, press and hold the AUX selector for approx. 2 seconds.

Please note: if desired, the operation can be changed so the T-coil can be activated by a brief press.



T-coil		
Status Front Description		
		Streamer on and ready
••••		When connected, you will hear a confirmation beep followed by the sound from the tele loop system


Steady

FM receiver (optional)

An FM receiver can be plugged into the Euro pin socket. The receiver is used to receive FM signals from an FM transmitter, typically worn by a school teacher.

Simply plug in the FM receiver and briefly press the AUX selector. When the FM transmitter and receiver are synchronised and on the same channel, the FM signal will be received in the hearing aids.

If your FM receiver has a toggle switch, position the switch onto the OO position.



IMPORTANT NOTICE

Please ensure the Streamer is turned on, but not streaming sound to the hearing aids, when FM channel sync is required.

Additional information

Cleaning and maintenance

The Streamer is a sealed device that requires minimum maintenance. Any dirt, moisture or grease should be cleaned off by using a soft, dry cloth. It is recommended that you clean the Streamer on a regular basis. The microphone inlet and the connectors for audio input, neck loop and charging must be kept free of dirt or debris.


IMPORTANT NOTICE

Use a soft, dry cloth to clean the Streamer. The Streamer and the neck loop must never be washed or immersed in water or other liquids.

Accessories and spare parts

Protective skins

Additional protective skins, dark or light grey, are available from your hearing care professional.



A protective skin, dark or light grey, with one protruding button, which makes "blind" operation easier, is available from your hearing care professional.





Replacement neck loops

Replacement neck loops, black or white, are available in three sizes: Short (51 cm), medium (66 cm) and long (81 cm) from your hearing care professional.





Belt clip

A belt clip, black or white, is available from your Hearing Care Professional.





Marnings

You should familiarise yourself fully with the following general warnings and the entire contents of this booklet before using your Streamer to ensure personal safety and correct use. Consult your hearing care professional if you experience unexpected operations or events with your Streamer.

Connection to external equipment

The safe use of the Streamer when connected to external equipment with an auxiliary input cable is determined by the external signal source. When connected to external equipment plugged into a wall outlet, this equipment must comply with IEC-62368 (or IEC-60065, IEC-60950 until June 20, 2019) or equivalent safety standards.

Rechargeable battery

Do not attempt to open the Streamer, as it may damage the battery.

Never attempt to replace the battery. If battery replacement is needed, please return your device to the supplier. The service guarantee is void if there are signs of tampering.

The safety of recharging batteries using a USB connector is determined by the external power source. When connected to external equipment plugged into a wall outlet, this equipment must comply with IEC-62368 (or IEC-60065, IEC-60950 until June 20, 2019) or equivalent safety standards.

Choking hazards and risk of swallowing batteries or other small parts

The Streamer, its parts and batteries should be kept out of reach of children and anyone who might swallow these items or otherwise cause injury to themselves.

If a battery, hearing aid or small part is swallowed, see a doctor immediately.

64

⚠ Warnings

Cross control

Your Streamer should be kept out of the reach of children or anyone who might take control of the functionality, to ensure your personal safety and avoid incorrect use.

Use of the Streamer with neck loop

If by accident the neck loop is caught on something while wearing it, the neck loop is constructed to break automatically. Therefore, never try to shorten or modify the neck loop in an unauthorised way. The neck loop should not be used for children who weigh less than 11 kg.

Heat and chemicals

The Streamer must never be exposed to extreme heat e.g. left inside a parked car in the sun. The Streamer must not be dried in microwave ovens or other ovens. The chemicals in cosmetics, hairspray, perfume, aftershave lotion, suntan lotion and insect repellent can damage the Streamer. Always remove your Streamer before applying such products and allow time to dry before use.

X-ray/CT/MR/PET scanning, electrotherapy and surgery

Remove your Streamer before X-ray/CT/MR/PET scannings, electrotherapy, surgery, etc. as your Streamer may be damaged when exposed to strong electromagnetic fields.

Active implants

Caution must be taken with active implants. In general, please follow the guidelines recommended by the manufacturers of implantable defibrillators and pacemakers on use with mobile phones and magnets as hearing aids and accessories also contain radio transmitters like mobile phones.

Marnings

Possible side effects

The otherwise non-allergenic materials used in your Streamer may in rare cases cause a skin irritation or other side effects.

Please seek consultation with a physician if these conditions occur.

Use in traffic

Using the Streamer can distract you. Be aware of your surroundings while using the Streamer.

Please refer to national traffic laws regarding the use of communication devices while in traffic.

Use on aircraft

Your Streamer contains Bluetooth. Your Streamer should not be used on board an aircraft, unless specifically permitted by the flight personnel.

Explosives

The power source in your Streamer has insufficient energy to cause fire in normal usage conditions. The Streamer has not been tested for compliance with International standards concerning explosive environments. We recommend not to use your Streamer in areas where there is a danger of explosions.

Interference

The Streamer has been thoroughly tested for interference according to the most stringent international standards. However, interference between the Streamer and other devices (e.g. some mobile telephones, citizens band devices, shop alarm systems and other wireless devices) may occur. If this occurs, increase the distance between the Streamer and the interfering device.

Warnings

Mobile phone compatibility

Please refer to www.oticon.global/connectline for a list of mobile phones that have been tested comprehensively by Oticon, and have been found to fulfil the minimum requirements for ConnectLine compatibility at the time of testing. However, Oticon takes no responsibility for such compatibility in general. Specifically, Oticon shall not be responsible for incompatibility resulting from future updates of the mobile phones, nor from events of incompatibility not covered by and/or found in these tests. Oticon takes no financial responsibility in terms of reimbursements of mobile phones that do not work as expected. Before purchasing a new mobile phone, you are advised to consult the list mentioned above.

Additiona

Technical information

Model name	Streamer Pro 1.3A		
Conditions of use			
Operating conditions	Temperature: +10°C to +45°C Humidity: <85% relative humidity Atmospheric pressure: 700 hPa to 1060 hPa		
Storage and transportation conditions	sportation extended periods during transportation and storage:		

The Streamer contains a radio transceiver using short range magnetic induction technology working at 3.84 MHz. The magnetic field strength of the transmitter is very weak and always below 15 nW (typically below -15 dBuA/m at 10 metre distance).

The Streamer contains a radio transceiver using Bluetooth at 2.4 GHz. The radio transmitter is weak and always below 10mW equal to 10 dBm in total radiated power.

The Streamer complies with international standards concerning radio transmitters, electromagnetic compatibility and human exposure.

Due to the limited space available on the Streamer relevant approval markings can be found in this booklet.

Additional information can be found in the: Technical Data Sheets (TD) on www.oticon.global.

USA and Canada

This device contains a radio module with the following certification ID numbers:

FCC ID: U28CL2STRM IC: 1350B-CL2STRM

Notice:

The device complies with Part 15 of the FCC Rules and with Industry Canada's licence-exempt RSSs standards. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

Notice:

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

This Class B digital apparatus complies with Canadian ICES-003.

NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules.

These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation.

If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment to an outlet on a circuit different from that to which the receiver is connected.
- Consult the manufacturer or an experienced radio/TV technician for help.

International warranty

Your Streamer is covered by an international limited warranty issued by the manufacturer for a period of 12 months from the date of delivery. This limited warranty covers manufacturing and material defects in the device itself, but not in accessories such as e.g. batteries, cables, power supplies, etc. Problems arising from improper/incorrect handling or care, excessive use, accidents, repairs made by an unauthorised party, exposure to corrosive conditions, damage due to foreign objects entering the device or incorrect adjustments are NOT covered by the limited warranty and may void it. The above warranty does not affect any legal rights that you might have under applicable national legislation governing sale of consumer goods.

If you need service

Take your Streamer to your hearing care professional, who may be able to sort out minor problems and adjustments immediately.

The manufacturer declares that this Streamer is in compliance with the essential requirements and other relevant provisions of Directive 2014/53/EU. Declaration of conformity is available at the manufacturer.



Manufactured by: Oticon A/S Kongebakken 9 DK-2765 Smørum Denmark

www.oticon.global









E Rinetod

Description of symbols used in this booklet



Warnings

Text marked with a warning symbol must be read before using the device.



Manufacturer

The device is produced by the manufacturer whose name and address are stated next to the symbol. Indicates the medical device manufacturer, as defined in EU Directives 90/385/EEC. 93/42/EEC and 98/79/EC.



CE mark

The device complies with Medical Device Directive 93/42/EEC.



Electronic waste (WEEE)

Recycle hearing aids, accessories or batteries according to local regulations. Hearing aid users can also return electronic waste to their hearing care professional for disposal. Electronic equipment covered by Directive 2012/19/EU on waste and electrical equipment (WEEE).



Regulatory Compliance Mark (RCM)

The device complies with electrical safety, EMC and radio spectrum requirements for devices supplied to the Australian or New Zealand market.



Indoor use


Indicates that the product is qualified for indoor use.



Double insulation

Indicates that product power supply is double insulated. This means a very low risk of single fault failures, i.e. single fault failures cannot occur.

74



Description of additional symbols used on labels



Keep dry

Indicates a medical device that needs to be protected from moisture.



Caution symbol

Consult instructions for use for warnings and cautions.



Catalogue number

Indicates the manufacturer's catalogue number so that the medical device can be identified.



Serial number

Indicates the manufacturer's serial number so that a specific medical device can be identified.



Power supply outlet

Indicates the rate at which power is sent from the outlet to your electrical device.



Consult instructions for use

Indicates the need for the user to consult an instructions for use.



Japan TELEC label and Japan JATE label
"R" indicates mark for certified radio equipment.
"T" indicates that the device can connect to the telephone network.

Singapore IMDA label



Indicates that the device complies with the standards and specification published by Info-Communications Media Development Authority (IMDA) and is compatible with the public telecommunication networks in Singapore and does not cause radio frequency interference to other authorized radio-communication networks.

Index	С	E	L
	Call	Earphones. See Headset	Landline phone 40
A	Answer 34	End the call 36	Last number redial 38
A2DP device. See Music wireless	End 34, 36	Euro pin socket. See FM	Listen to audio 48
Accessories 62	Make 36		Listen to music 48
Aircraft, use on 66	Reject 38	F	M
Answer a call 34	Transfer 39	First-time use 12	Maintenance 61
Antenna 12	Changing programs 24	FM 58	Make a call 36
Audio (music) 48	Charging 18		Microphone 42
Cable 48	Cleaning 61	Н	Microphone inlet 11
Listen to music 48	Clear pairing 28	Headphone. See Headset	Mini jack 11
Wireless 52	Computer	Headset 9	Mini jack cable 48, 50
	Headset cable 50	Cable 50	Mobile phone 33
В	Music/audio cable 48	Mobile phone 33	MP3 player 52
Battery 22	Music/audio wireless 52	Splitter cable 50	Music 52
Capacity 22	ConnectLine	Wireless 52, 54	Cable 48
Charge time 22	Phone 40	Headset cable 50	Listen to music 48
Life 22	TV 44	Headset splitter cable 50	Wireless 52
Low 20			Mute 24
Rechargeable 18		I	Muting the hearing instruments 24
Bluetooth 52		Instrument view 11	
	78		79

N Neck loop 12	S Sound, See music
Neck loop 12	Spare parts 62
P	Splitter 9
Pairing 26	Splitter cable 50
Clearing 28	Static electricity 30
Mode 26	
PC. <i>See</i> Computer	Т
Personalising 38	Talking on the phone 41
Phone solution 40	T-coil 56
Pin code 27	Technical information 68
Power off/on 16	Tele loop 56
Protective skin 62	Transfer call 39
	TV 44
R	
Rechargeable battery 18	V
Reject call 38	Voice dial 38
Remote control 24	Volume up/down 24

W Warnings 64, 65, 66, 67 Wireless 52, 54 Without the neck loop 14



