




**UltraVac+**  
Vacuum & Drying Chamber

## Hearing Instrument Restoration System

Designed to give dispensers & technicians a functional tool for repairing, restoring and maintaining hearing aids.

Effectively removes cerumen and debris from hearing aid tubing and ports.



# UltraVac+

## Technical Specs

**Power Requirements:** 100 – 240 V, 50/60 Hz

**Power Consumption:** <100 Watts

**Fuse:** 2.5A @ 100-240V

**Operating Temperature:** 0° To 50° C • 32° To 120° F

**Debris Filter:** I-63S

**Vacuum Tips:** #14, #15, 2-#18 and 2-#20 Gauge With Safety Lock Attachment

**Drying Chamber Cycle Time:** 4 Minutes With Auto Shut-off

**Vacuum Wand:** Continuous Use

**Pressure Wand:** Continuous Use

**Dimensions:** Approx. 29 cm x 31 cm x 16 cm (L x W x H)  
Approx. 11.25" x 12" x 6.25" (L x W x H)

**Weight:** Approx. 5 kg • Approx. 11 lbs

**Standard Accessories:** Drying Chamber Cup With Filter, Vacuum Wand With Filter, Pressure Wand, 6 Vacuum Wand Tips, Cleaning Tool Kit, Instruction Manual

## MedRx ULTRA VAC

- Digital Display of Chamber Pressure and Time Remaining
- Automatic Detection of Operating Mode
- Separate Pressure and Vacuum Wands Minimize Debris in System
- In-Line Vacuum Filter for Added Reliability
- 4 Minute Drying Chamber Cycle
- Auto Shut Off

### Functions of your UltraVac+

**The Drying Chamber** – Displaces moisture trapped within the hearing instruments. You can place two ITEs, or one BTE with an earmold, into the chamber at one time.

**The Vacuum Wand** – Used to extract ear wax and debris from the receiver tubes, microphone tubes, vents and battery compartments of the hearing instruments.

**The Pressure Function** – Used to clear obstructed vacuum wand tips and to blow debris from the receiver tubes, microphone tubes, vents and battery compartments of the hearing instruments.

# MedRx®

Good Things Come in Small Packages

### MedRx International

MAICO Diagnostics GmbH

Sickingenstr. 70-71

10553 Berlin, Germany

Phone: +49 30 70 71 46 50

Fax: +49 30 70 71 46 99

Email: [medrx-sales@maico.biz](mailto:medrx-sales@maico.biz)

[www.medrx-diagnostics.com](http://www.medrx-diagnostics.com)

Australian Distributor

**SONIC**  
EQUIPMENT

1800 639 263  
[info@soniceq.com](mailto:info@soniceq.com)  
[soniceq.com](http://soniceq.com)



The Vacuum & Pressure Wands work simultaneously.

